

湖北省计量测试技术研究院

Huber Institute of Measurement and Testing Technology

地址 (Add) : 湖北省武汉市东湖新技术开发区茅 邮编 (Post Code) : 430223
店山中路二号 电话 (Tel) : 027-81925136
网址 (Web site) : http://www.hbjl.gov.cn 传真 (Fax) : 027-81925137

中国校准服务联合体
China United Calibration Service

CUC

REGISTRATION NO. 002

校 准 证 书

CALIBRATION CERTIFICATE

证书编号: []
Certificate No.

委托方名称 Customer 中铁电气化局集团有限公司沈阳电气化工程分公司
委托方地址 Address 辽宁省沈阳市浑南区上深沟村861-19号 (231)
样品名称 Name Of Sample 回路电阻测试仪
制造厂商 Manufacturer 武汉特高压电力科技有限公司
型号规格 Model/Type UHV-H100P
器具编号 No Of Sample 1710203

证书专用章
Stamp



校准日期 Calibration date

2022年 09月 29日
Y M D

证书批准人
Approved by

核验员
Checked by

校准员
Calibrated by

耿强
王强

本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基标准及国家计量研究院出具的校准和测量证书相互承认协议”。

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.



湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology

证书编号: [2021]000000000000 号

Certificate No.

- 本院是政府计量行政管理部门依法设立的法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

- 本院质量管理体系符合 ISO/IEC17025 标准的要求。

The quality management system for laboratory complies with ISO/IEC 17025 standards.

- 本次校准的技术依据(名称、代号)

Reference documents for the Calibration (Name、Code)

参照: JJG1052—2009 回路电阻测试仪、直阻仪检定规程 Calibration for Fibre Fineness Tester
of Air-flow Method

- 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

设备名称

Name of Equipment

模拟大功率交直流标准电阻器

型号/编号

Model/Serial No.

MJZ-600/12037

证书号/有效期

Certificate No./Due Date

2022DW02250493/2023-09-30

- 校准环境条件

Environmental condition on the Calibration

温度: 21.0°C

相对湿度: 66 %

其它: _____

Temperature

R.H.

Others

气压: _____

地点: 本院光谷基地B211室

原始记录编号: 2022DW02250493

Pressure

Place

Record No.

本校准结论,仅对受校样品的本次校准有效。

It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated.

未经本院许可,不得部分复制本证书。

校准数据/结果

Data/Results of Calibration

标准值 ($\mu\Omega$)	被校显示值 ($\mu\Omega$)	测量不确定度
100.0	99.90	$U_{rel}=0.2\%, k=2$
200.0	200.0	
300.0	300.0	
400.0	400.0	
500.0	500.1	
600.0	600.3	
700.0	700.7	
800.0	800.8	
900.0	900.6	
1000	1000.7	

注: 校准时电流为 100A
以下空白

研究院